



MEETING TO FINALIZE THE SPECIFICATIONS OF THE VETERINARY EQUIPMENT
No. DAH-DPC-02/2017-18
Departmental Purchase Committee, Directorate of Animal Husbandry, 2nd Floor, Block-B,
Krishi Bhavan, Sector 10-A, GANDHINAGAR - 382010,
GUJARAT STATE. Phone : 079-23256138, Fax : 079-23256142

A Meeting is organized to finalize generic specifications for following Equipment in the Conference Hall of this office. Manufacturers, Direct Importers, Subsidiary of Original Manufacturers, Dealers and Distributors of this equipment are invited to remain present in this meeting. Only one representative from one company shall be allowed to participate in this meeting. Each participant shall be required to submit (1) Company's Authorization Letter, (2) Valid Identity Card and (3) Product Literature mentioning that the company manufactures this item.

Sr. No.	Name of Equipment	Date and Time
1.	Suction Pump	12-09-2017, 12:00 hrs
2.	Metal Detector for Veterinary Use	
3.	Veterinary Ultrasound Machine	
4.	Electrosurgical Unit	
5.	Electric ULV Fogging Machine	

The specifications for above equipments are available on Department's website: <https://doah.gujarat.gov.in/>, which may be downloaded free of cost. The same is also available on Notice Board of this office.

Date : 06-09-2017
Place : Gandhinagar

Chairman, Departmental Purchase Committee cum
Director of Animal Husbandry, Gujarat State,
Gandhinagar



વેટરનરી ઇક્વિપમેન્ટનાં સ્પેસીફિકેશન્સ આખરી કરવા અંગેની મીટીંગ
નં. DAH-DPC-02/2017-18

ખાતાકીય ખરીદ સમિતી, પશુપાલન નિયામકની કચેરી, બીજો માળ, બી-બ્લોક, કૃષિ ભવન, સેક્ટર-૧૦એ,
ગાંધીનગર - ૩૮૨૦૧૦, ગુજરાત રાજ્ય. ફોન : ૦૭૯-૨૩૨૫૬૧૩૮, ફેક્સ : ૦૭૯-૨૩૨૫૬૧૪૨

નીચે દર્શાવેલ સાધનના સ્પેસીફિકેશન્સ આખરી કરવા માટેની મીટીંગ આ કચેરીના 'કોન્ફરન્સ હોલ' ખાતે રાખવામાં આવેલ છે. આ સાધનના ઉત્પાદકો, ડાયરેક્ટ ઇમ્પોર્ટર્સ, મૂળ ઉત્પાદક કંપનીની પેટા કંપનીઓ, ડીલર્સ અને વિતરકોને મીટીંગમાં ઉપસ્થિત રહેવા આમંત્રિત કરવામાં આવે છે. દરેક કંપની વતી ફક્ત એક જ પ્રતિનિધિ આ મીટીંગમાં હાજરી આપી શકશે અને મીટીંગમાં હાજર રહેનાર પ્રતિનિધિએ (૧) કંપનીનો ઓથોરાઈઝેશન લેટર, (૨)માન્ય ઓળખપત્ર અને (૩)કંપની આ ઓર્ડરમનું ઉત્પાદન કરે છે તે દર્શાવતું પ્રોડક્ટ લીટરેચર રજૂ કરવાનું રહેશે.

અ.નં.	સાધનનું નામ	તારીખ અને સમય
૧.	સકશન પંપ	૧૨-૦૯-૨૦૧૭, ૧૨:૦૦ કલાકે
૨.	મેટલ ડિટેક્ટર ફોર વેટરનરી યુઝ	
૩.	વેટરનરી અલ્ટ્રાસાઉન્ડ મશીન	
૪.	ઇલેક્ટ્રોસર્જીકલ યુનિટ	
૫.	ઇલેક્ટ્રિક યુ.એલ.વી. ફોર્ગીંગ મશીન	

સદરહું સાધનના સ્પેસીફિકેશન્સ ખાતાની વેબસાઈટ: <https://doah.gujarat.gov.in/> પરથી વિના મુલ્ય મેળવી શકાશે, સ્પેસીફિકેશન્સ આ કચેરીના નોટીસ બોર્ડ ઉપર પણ ઉપલબ્ધ છે.

તારીખ: ૦૬-૦૯-૨૦૧૭
સ્થળ: ગાંધીનગર

અધ્યક્ષ ખાતાકીય ખરીદ સમિતિ
સહ
પશુપાલન નિયામક, ગુજરાત રાજ્ય, ગાંધીનગર

TENTATIVE SPECIFICATIONS OF SPECIAL EQUIPMENTS - 2017-18

Sr. No.	Name of the Item	Tentative Specifications
1	2	3
1	SUCTION PUMP	<ul style="list-style-type: none"> • High quality low noise electric suction units choice of power options : mains of mains plus battery collection jar fittings are compatible with other popular canister/liner systems integrated overflow valve oil –free, No maintenance pump • Bacteria filter and suction house V7M is supplied with a 12 / 24 V power cable for in-vehicle use. Must have foot switch. • Capacity for large animal.
2	METAL DETECTOR FOR VETERINARY USE	<ul style="list-style-type: none"> • Detecting covered Magnetic and nonmagnetic Metal parts. • FED 3 has optical notice on scale 0-30 • Additionally on acoustic tone via earphone. • Ability of scale marking for localizing of a foreign body. • Instrument consists of Metal detector instrument with probe, earphone batteries.
3	ELECTROSURGICAL UNIT	<ul style="list-style-type: none"> • Solid state electrosurgical. • Generator with 250 W cut, 100 W coagulation and 80 W bipolar mode. • Operating frequency: 480 KHz with Standard accessories. • Chuck handle with electrodes autoclavable Bipolar forceps with cable autoclavable : Monopolar double pedal foot switch / Bipolar single pedal foot switch • Disposable hand switch pencil. • Patient plat with cord autoclavable.

4	VET ULTRA SOUND MACHINE	<ul style="list-style-type: none"> • (ENDO RECTAL PROBE + BATTERY , ALONG WITH FOOTSWITCH, MOBILE TROLLEY, CARRYING CASE,) • Rectal Probe : frequency 3 MHZ to 8MHZ • Built in battery (shall provide minimum 60 minits back up) • Foot switch • Mobile Trolley • Carrying case • 15.00" high resolution LCD screen • Machine weight : maximum 10 kg • Color Doppler Flow Imaging • Pulsed wave Doppler • Directional Power Doppler Imaging. • Tissue Harmonic Imaging • Compound Imaging • Speckle Imaging • Boot up Time : Shall not be more than 2 mints • The machine shall perform more than 200 frames per second
5	Electric ULV Fogging Machine:	<ul style="list-style-type: none"> • Must be ideal for aerosol spraying of aqueous disinfectant in Hospital, animal/poultry houses, Dairy Unit etc. • Droplet size generated must be very small in microns (not >50 microns) • Must be portable in design • With timer • Must have nozzle or Valve for fog output control • Input Power 230 Volts Single Phase AC, 50/60 Hz. • Tank Capacity Range : 5 to 8 litre • Tank must be made up of HDPE • Body parts must resist corrosion or chemical attack • Motor Speed should be 20,000 RPM (approx.) or more • Area Coverage: Must be more than 6000 cubic fts